en

Instruction for use (Dental Absorbent Paper Points)

*hand rolled for extra absorbency
- Highly absorbent
- Length marked available
- Variety sizes and packages available
- Color coded

INDICATION AND PURPOSE

Paper points are used to dry the root canal before obturation and to clean the root canals that are to be sealed in endodontic treatments.

INSTRUCTION FOR USE

1. Measure the length of root canal
2. The formation of the root canal of the same size as the last file to select
3. The root canal using paper points to the dry canal
4. About 10 – 15 points is used per a root canal
5. Sealer or cement is applied by using paper points

CAUTIONS

1. Should be use sterilized paper points
2. Do not use bent or broken points
3. Should be careful to avoid overfilling or dead space when applied sealer or cement
4. Use it immediately when applied sealer or cement
5. After using the paper points, should be kept in a cool place
6. Do not expose to direct sunlight
7. Do not reuse to prevent cross infections. (Secondary infections such as pathogens and viruses caused by saliva and blood)

STORAGE

Store at room temperature of 1°C ~ 30°C(33°F ~86°F)

SHELF LIFE

5 years form manufacture date

AVAILABLE SIZE

Type	Sizes
Standardized (ISO Specifications)	10, 15, 20, 25, 30, 35, 40, 45, 50, 55, 60, 70, 80, 90, 100, 110, 120, 130, 140
Non-Standardized (Accessory Sizes)	XX-Fine, X-Fine, Fine, Medium, Coarse, X-Coarse, Protaper Size(F1~F5)
Special Taper (.04, .06, .08)	15, 20, 25, 30, 35, 40, 45, 50 55, 60, 70, 80

COMPONENTS

- **Available packaging**
- 1, 4 compartment dispenser – Code: P21/Regular packing points: 60/200 points
- 2, Sliding type dispenser – Code: P71/Regular packing points: 60/200 points
- 3, Vial packing(Sliding type) – Code: P31/Regular packing points: 60/200 points
- 4, Vial packing(Round) – Code: P31/Regular packing points: 60/200 points
- 5, 4 compartment dispenser – Code: P51/Regular packing points: 200 points
- 6, Blister Cell – Code: P61/Regular packing points: 180/200 points

Symbols

	Use By
LOT	Batch Code
STERILE R	Sterilization using irradiation
EC REP	Authorized representative in the EC
	Consult Instruction For Use
	Do not reuse
	Do not irradiate
	Manufacturer
	Date of manufacture
	Keep away from sunlight
	Temperature limitation
CE NB	CE – Mark and identification number of notified body. Product conforms to the essential requirements of the Medical Device Directive 93/42/EEC

el

Οδηγίες Χρήσης (Οδοντικοί Απορροφητικοί Κώνοι Χάρτου)

Διπλωμένοι "χειρί" για καλύτερο απορροφητικό
- Υψηλής απορροφητικότητας
- Διαθέσιμα ανιχνευτήρια
- Διαθέσιμα σε διάφορα μεγέθη και πακετα
- Χρωματικοποιημένα

ΕΝΔΙΑΦΡΟΝΗ

Οι κώνοι χαρτιού χρησιμοποιούνται για την ξήρανση του ριζικού σωλήνα πριν από την σφραγίδα και για τον καθαρισμό των ριζικών σωλήνων που πρόκειται να σφραγιστούν στη ενδοδοντικές θεραπείες

ΟΔΗΓΙΕΣ ΧΡΗΣΗΣ

1. Μετρήστε το μήκος του ριζικού σωλήνα
2. Επιλέξτε τον κώνο χαρτιού που ταιριάζει με το μέγεθος της τελευταίας λίμας που χρησιμοποιήθηκε για τον ριζικό σωλήνα.
3. Χρησιμοποιήστε τους κώνους χαρτιού για να στεγνώσετε τον ριζικό σωλήνα.
4. Χρησιμοποιούνται περίπου 10 - 15 κώνοι ανά ριζικό σωλήνα
5. Η σφραγιστική αφροκολλητική ή τσιμεντένια ενίεται με κώνους χαρτιού

Προσοχή

1. Απαιτείται η χρήση αποστειρωμένων κώνων χαρτιού
2. Μην χρησιμοποιείτε λυγισμένους ή σπασμένους κώνους
3. Τηρείτε και εξυπνάτε προσεκτικά κάθε φορά που εφαρμόζετε ή τη νοκρό θιάστημα στην εφαρμόζετε αφροκολλητική ή τσιμέντο
4. Χρησιμοποιήστε το αμέσως στην εφαρμόζετε αφροκολλητική ή τσιμέντο
5. Οι κώνοι χαρτιού πρέπει να φυλασσόνται σε δροσερό μέρος μετά τη χρήση
6. Μην εκθέτετε σε άμεσο ηλιακό φ
7. Μην επαναχρησιμοποιείτε κώνο για πρόληψη διάσπασης ή ρύπανσης (Δευτερογενείς λοιμώξεις: ασθενείς πρόκαυσης και του ιού ιού ιού

품 명	PP 사용설명서
서식번호	DM-124-4
치 수	698X222
용지재질	모조 45g
작 업 자	홍선미
Printing Color	PANTONE 540C

Revised Date	Explications
2016. 03. 21	신규등록
2017. 05. 29	DM-124-1 / 내용 수정
2018. 11. 14	DM-124-2 / CI.심볼 수정, EC REP 정보변경
2020. 02. 17	DM-124-3 / CE번호 변경
2020. 10. 10	DM-124-4 / 다국어 적용