

Smartcord X

Description
Smartcord X is a knitted retraction cord, made of 100% cotton, impregnated with Aluminum Chloride Hexahydrate for temporary gingival retraction and hemostasis of the gingival margin.

Indications
Smartcord X is used to aid in making impressions for fixed prosthetics and is also used for gingival displacement and hemostasis before bonding and/or restorative placement procedures.

Options
Smartcord X is available in 6 diameters identified by different colors.

| | |
|-----------------------------|---------------------|
| Smartcord X 04 (The Finest) | black/blue |
| Smartcord X 03 (Ultra-Fine) | black/gray |
| Smartcord X 02 (Extra-Fine) | brown/yellow |
| Smartcord X 0 (Fine) | purple/light purple |
| Smartcord X 1 (Medium) | blue/light blue |

Smartcord X 2 (Thick) Do not re-use
green/light green

Composition
100% Cotton yarn 254cm(100inch) impregnated with Aluminum Chloride Hexahydrate 1.2±0.8mg per inch of cord.

Procedure
1. Select a cord with an appropriate diameter.
2. Cut the cord to the required length with EasyCut and SmartRuler. If necessary, the cord can be impregnated with a hemostatic solution.
3. Insert the cord into the sulcus by means of a retraction cord packer.
4. Leave the cord about 5 minutes in the sulcus.
5. Remove the cord just before taking the impression. Slightly moisten the cord to make its removal easier and reduce gingival hemorrhage.

Precautions
- Check the conditions of gingival tissue(eg. depth or thickness)
- For professional use only.

- For single use only.
- To avoid a possible permanent displacement of the gingiva, the cord should not be remained for longer than 20 minutes in the sulcus.
- After finishing the procedure, make sure to remove the cord from the sulcus.
- The package must be close immediately after use.
- Store in a dry, dark place between 1 °C and 30 °C
- Shelf life is 3 years. After opening, the bottle may be used until the expiration date.

Contraindications
⚠ Do not use in patients with an allergy to Aluminum Chloride Hexahydrate.
Smartcord X should not be used in case of acute gingivitis or acute necrotizing ulceric gingivitis. The use of Smartcord X in the inflamed area should be considered carefully, weighing the benefits against the risks, as the retraction of the gingival margin could cause tissue damage a certain extent.

Symbols

CE mark **MD** Medical Device Do not re-use Caution Keep dry Use-by date Keep away from sunlight

For professional use only Temperature limit Manufacturer Consult instructions for use LOT Batch code Authorized representative in the European Community

Manufacture **EASTDENT CO., LTD**
64, Gugaj-ro, Giheung-gu, Yongin-si, Gyeonggi-do, 16972, Korea
Tel : +82-31-693-5767, Fax : +82-31-693-5767

KTR Europe GmbH
Mergenthalerallee 77, 65760, Eschborn, Germany
Tel: +49-6196-887-170
EDXIE, Rev4, 03/2022

Smartcord X

Description
Smartcord X is a knitted retraction cord, made of 100% cotton, impregnated with Aluminum Chloride Hexahydrate for temporary gingival retraction and hemostasis of the gingival margin.

Indications
Smartcord X is used to aid in making impressions for fixed prosthetics and is also used for gingival displacement and hemostasis before bonding and/or restorative placement procedures.

Options
Smartcord X is available in 6 diameters identified by different colors.

| | |
|-----------------------------|---------------------|
| Smartcord X 04 (The Finest) | black/blue |
| Smartcord X 03 (Ultra-Fine) | black/gray |
| Smartcord X 02 (Extra-Fine) | brown/yellow |
| Smartcord X 0 (Fine) | purple/light purple |
| Smartcord X 1 (Medium) | blue/light blue |

Smartcord X 2 (Thick) Do not re-use
green/light green

Composition
100% Cotton yarn 254cm(100inch) impregnated with Aluminum Chloride Hexahydrate 1.2±0.8mg per inch of cord.

Procedure
1. Select a cord with an appropriate diameter.
2. Cut the cord to the required length with EasyCut and SmartRuler. If necessary, the cord can be impregnated with a hemostatic solution.
3. Insert the cord into the sulcus by means of a retraction cord packer.
4. Leave the cord about 5 minutes in the sulcus.
5. Remove the cord just before taking the impression. Slightly moisten the cord to make its removal easier and reduce gingival hemorrhage.

Precautions
- Check the conditions of gingival tissue(eg. depth or thickness)
- For professional use only.

- For single use only.
- To avoid a possible permanent displacement of the gingiva, the cord should not be remained for longer than 20 minutes in the sulcus.
- After finishing the procedure, make sure to remove the cord from the sulcus.
- The package must be close immediately after use.
- Store in a dry, dark place between 1 °C and 30 °C
- Shelf life is 3 years. After opening, the bottle may be used until the expiration date.

Contraindications
⚠ Do not use in patients with an allergy to Aluminum Chloride Hexahydrate.
Smartcord X should not be used in case of acute gingivitis or acute necrotizing ulceric gingivitis. The use of Smartcord X in the inflamed area should be considered carefully, weighing the benefits against the risks, as the retraction of the gingival margin could cause tissue damage a certain extent.

Symbols

CE mark **MD** Medical Device Do not re-use Caution Keep dry Use-by date Keep away from sunlight

For professional use only Temperature limit Manufacturer Consult instructions for use LOT Batch code Authorized representative in the European Community

Manufacture **EASTDENT CO., LTD**
64, Gugaj-ro, Giheung-gu, Yongin-si, Gyeonggi-do, 16972, Korea
Tel : +82-31-693-5767, Fax : +82-31-693-5767

KTR Europe GmbH
Mergenthalerallee 77, 65760, Eschborn, Germany
Tel: +49-6196-887-170
EDXIE, Rev4, 03/2022